



# OKmeter Cameo

## Blood Glucose Monitoring System



*It's time to  
travel light*

- **No Coding Required**
- **Alternative Site Testing**
- **Mini Size Meter**

Accurate results in 6 seconds


Very small blood sample, only 0.7uL

120 test memory with date and time

5 years meter warranty



## STEPS



# 1

Insert test strip to turn on meter.

# 2



Apply small blood sample.



# 3

Obtain results in 6 seconds.



ISO13485 CE0123

# SYSTEM ACCURACY (ISO 15197)

**Test system:** OKmeter  
**Reference system:** YSI2300  
**Sample:** Whole blood (capillary)

**Investigative Site:**  
 Institut für Diabetes-Technologie GmbH an der  
 Universität Ulm (IDT)  
 Helmholtzstrasse 20, D-89081 Ulm / Germany


## System accuracy results for glucose concentration < 4.2 mmol/L (75 mg/dL)

Within ± 0.28 mmol/L (Within ± 5 mg/dl)	Within ± 0.56 mmol/L (Within ± 10 mg/dl)	<b>Within ± 0.83 mmol/L</b> (Within ± 15 mg/dl)
19 / 40 (48 %)	34 / 40 (85 %)	<b>40 / 40 (100 %)</b>

## System accuracy results for glucose concentration > = 4.2 mmol/L (75 mg/dL)

Within ± 5 %	Within ± 10 %	Within ± 15 %	<b>Within ± 20 %</b>
64 / 160 (40 %)	118 / 160 (74 %)	154 / 160 (96 %)	<b>159 / 160 (99 %)</b>

The OKmeter blood glucose system complies with the system accuracy requirements of the ISO 15197 Standard. 199 of 200 (99.5%) results meet the requirements.



# SPECIFICATIONS

Model Name	OK-9P
Assay Method	Electrochemical biosensor
Test Sample	Capillary Whole Blood
Test Result	Referenced to plasma glucose value
Alternative Site Testing	YES (palm, forearm, upper arm, calf, and thigh)
Sample Size	0.7 µL
Measuring Time	6 seconds
Measuring Range	20 – 600 mg/dL (1.1 – 33.3 mmol/L)
Acceptable HCT Range	20~60%
Operating Condition	10°C~40°C (50~104°F) / 10 ~ 85% R.H. Altitude ≤ 3,402 m (11,161 ft) above the sea level
Storage/Transportation Condition	4~40°C (39~104°F) 10~85% R.H.
Memory Capacity	120 test results with time and date
Average Capacity	7, 14, and 28 days
Power Supply	One 3-volt Lithium Battery (battery type CR2032)
Battery Life	Approximately 1,000 tests
Dimensions	90mm(L) × 35mm(W) × 21mm(H)
Weight	37 g



Manufacturer

**OK Biotech Co., Ltd.**

4F-1, No. 83, Sec. 2, Gongdao 5th Rd.,

300 Hsinchu City, Taiwan

TEL: +886-3-5160258 FAX: +886-3-5160028

<http://www.okbiotech.com>

<http://www.glucose-meter-manufacturer.com>

[service@okbiotech.com](mailto:service@okbiotech.com)



OK-9P-EN-120522